

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/09/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:38
INTERNAL STANDARD	VERIFIED	01:39
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42



SGT Ralph J Bartlett #162 12-09-09