

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:49
INTERNAL STANDARD	VERIFIED	00:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:52



SOB Ralph Bartlett weso #162 12/17/09