

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/25/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:11
INTERNAL STANDARD	VERIFIED	21:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	21:12
BLANK TEST	0.00	21:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	21:13
BLANK TEST	0.00	21:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:14
BLANK TEST	0.00	21:15



SGT Ralph Bartlett 12-25-09 WESO SHERIFF #162