

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 02/13/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:07
INTERNAL STANDARD	VERIFIED	00:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:08
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10



SGT *Ralph Bartlett*

2-13-09

SHERIFF #162