

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 02/17/2009

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:22
INTERNAL STANDARD	VERIFIED	02:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25



SG *Ralph Bartlett* SHERIFF # 162