

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/23/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:16
INTERNAL STANDARD	VERIFIED	03:16
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:19
BLANK TEST	0.00	03:19



SGT Ralph J Bartlett

SHERIFF # 102