

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/03/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:31
INTERNAL STANDARD	VERIFIED	02:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:34
BLANK TEST	0.00	02:35



SGT *Ralph Bartlett* WCSO # 162