

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:34
INTERNAL STANDARD	VERIFIED	01:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:35
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37



S65 *Ralph Bartlett* Weso SHELDON #162