

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/23/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 08370

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:32
INTERNAL STANDARD	VERIFIED	04:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:34
BLANK TEST	0.00	04:35



507 *Ralph Bartlett* WCSO Sheriff #162