

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/01/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34



SOA Rusky ~~Buss~~ WESO SHERIFF #162