

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/08/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:45
INTERNAL STANDARD	VERIFIED	17:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:46
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:48



*See Ralph Bartlett WCSO #162*