

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/15/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:22
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:23
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25



SGS Ralph J Bartlett

WCSO #162