

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/23/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:57
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:57
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:58
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:59
BLANK TEST	0.00	02:00



SGS Ralph Bartlett

weso #162

4/23/09