

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/07/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34



SGT Ralph J Bartlett WPSO # 162