

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/13/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:53
INTERNAL STANDARD	VERIFIED	03:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:55
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57



565 *Ralph Bartlett*

SHELDON # 142