

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/27/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:59
INTERNAL STANDARD	VERIFIED	00:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:02



SGT Rupp / Review #112