

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/02/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:27
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29



SGT *Ralph Bartlett*

SAT #102

7/2/09