

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/08/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:52
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:53
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:54
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:55
BLANK TEST	0.00	00:55



SGT Ralph Bartlett SHERIFF #162 7/8/09