

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 07/16/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:39
INTERNAL STANDARD	VERIFIED	03:39
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:40
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:42
BLANK TEST	0.00	03:42



SO Ralph Bartlett

WCSO #162 7/16/09