

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/23/2009

Operator Name: KENNETH W SMITH  
Permit #: 35604  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:34
INTERNAL STANDARD	VERIFIED	02:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:35
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:37
BLANK TEST	0.00	02:37



*Handwritten signature* # 301

07/23/2009