

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/29/2009

Operator Name: KENNETH W SMITH
Permit #: 35604
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:42
INTERNAL STANDARD	VERIFIED	01:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:44
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	01:45
BLANK TEST	0.00	01:45



KWS #301

7/29/2009