

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/04/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:33
INTERNAL STANDARD	VERIFIED	02:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:34
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:35
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:36



*SOS Rush / [Signature]*

*WCSO #162*