

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 08/19/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:44



SCS Ralph J Bartlett WCSO SHERIFF # 162 8/19/09