

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/27/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:39
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:40
REFERENCE STANDARD TEMP	34.0c	
BLANK TEST	0.00	02:41



SO5 Ralph J Bartlett 8/27/09 Sheriff # 162