

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/02/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09150

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 02:16 |
| INTERNAL STANDARD | VERIFIED | 02:16 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 02:16 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 02:17 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 02:18 |
| BLANK TEST | 0.00 | 02:19 |



SCS *Ralph Bartlett*

weso #162 9/2/09