

SUPERVISOR TEST




WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 09/24/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09150

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:45
INTERNAL STANDARD	VERIFIED	02:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:45
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:46
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:47
BLANK TEST	0.00	02:48



SOS Ralph Bartlett

9/24/09