

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 09/30/2009

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:05
INTERNAL STANDARD	VERIFIED	03:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08



*SG Ralph Bartlett* WARREN CO. SHERIFF'S OFFICE #162 9/30/09