

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/08/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:09
INTERNAL STANDARD	VERIFIED	02:10
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:10
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:11
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:12
BLANK TEST	0.00	02:13



SOE Ralph Bartlett

10/8/09