

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/17/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:47
INTERNAL STANDARD	VERIFIED	03:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:47
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:48
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50



SOS Ralph Bartlett

WCSO #162

10/17/09