

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/17/2009

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09200

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:05
INTERNAL STANDARD	VERIFIED	10:05
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	10:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	10:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	10:08
BLANK TEST	0.00	10:08

Charles B. Geisler #248

