

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 12/09/2009

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:59
INTERNAL STANDARD	VERIFIED	01:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02



SGS *Ralph Bartlett* WCSO #162 12/9/09