

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/17/2009

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09260

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 01:08 |
| INTERNAL STANDARD       | VERIFIED | 01:08 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 01:08 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 01:09 |
| REFERENCE STANDARD TEMP | 34.0c    |       |
| REFERENCE STANDARD      | 0.09     | 01:10 |
| BLANK TEST              | 0.00     | 01:11 |



SOT *Ralph Bartlett* WCSO SHERIFF #162 12-17-09