

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/25/2009

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 20:54 |
| INTERNAL STANDARD | VERIFIED | 20:54 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:55 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:56 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 20:57 |
| BLANK TEST | 0.00 | 20:58 |



SGT Ralph J Bartlett WCSO #142 12/25/09