

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/08/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:44
INTERNAL STANDARD	VERIFIED	07:44
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	07:45
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	07:46
BLANK TEST	0.00	07:47



SOB Ralph Bartlett

WASO #162

1/4/10