

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:16
INTERNAL STANDARD	VERIFIED	03:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:16
BLANK TEST	0.00	03:17
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:17
BLANK TEST	0.00	03:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:18
BLANK TEST	0.00	03:19



SGT Ralph Bartlett

WCSO #162

1/14/10