

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/20/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:57
INTERNAL STANDARD	VERIFIED	02:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01



505 *Ralph Bartlett* vcs0 #162

1/20/10