

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/17/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:56
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:57
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:58
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00



S65 Rugh / Burt 2-17-10 W050 #162