

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/25/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:47
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:50



SGT Ralph Bartlett

02-25-10

WISO #162