

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/08/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:56
INTERNAL STANDARD	VERIFIED	03:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:59



SGT *Ralph Bartlett*

WASO #142

3-8-10