

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/16/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:55
INTERNAL STANDARD	VERIFIED	03:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:57
BLANK TEST	0.00	03:58



505 *Ralph Bartlett*

WCSO #162

3-16-10