

SUPERVISOR TEST

-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 105208  
 -----

Date: 04/07/2010

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:00
INTERNAL STANDARD	VERIFIED	00:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:02
BLANK TEST	0.00	00:03

