

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58



SUT Ralph Bartlett

WISO #162

04-14-10