

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/22/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:15
INTERNAL STANDARD	VERIFIED	22:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:17
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:18
BLANK TEST	0.00	22:18



SG Ralph Bartlett w/ISO #162 4/22/10