

SOLUTION CHANGE

 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 105208

Date: 05/02/2010

Operator Name: CHARLES B GEISLER
 Permit #: 33757
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:54
INTERNAL STANDARD	VERIFIED	08:54
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	08:54
BLANK TEST	0.00	08:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:55
BLANK TEST	0.00	08:56
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	08:56
BLANK TEST	0.00	08:57

Charles B. Geisler #33757

