

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/18/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:35
INTERNAL STANDARD	VERIFIED	02:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38



505 Ralph Bartlett 06/18/10