

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/26/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:43
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:44
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:45
BLANK TEST	0.00	23:45



505 Rugby Buses

06/27/10