

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/15/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:42
INTERNAL STANDARD	VERIFIED	17:42
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:43
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46



SGT Rusky ~~T. Burt~~ WESO #162 7/15/10