

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/09/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:09
INTERNAL STANDARD	VERIFIED	01:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:12



SGS Rush/Budget WC30 #162 8/9/10