

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/11/2010

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:55
INTERNAL STANDARD	VERIFIED	14:55
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	14:55
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:58
BLANK TEST	0.00	14:59

J.M.L. #299

