

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/17/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:15
INTERNAL STANDARD	VERIFIED	01:15
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19



SGT *Ralph Bartlett*

WARREN CO SHERIFF #162

8-17-10